

DIAMOND[®]
D I A G N O S T I C S



CARELYTE[®]

The Choice for Veterinary Laboratories



CARELYTE®



SPECIFICATIONS:

Sample

Feline, Canine, Bovine, Equine, Swine, Ovine, Open

Sample Type

Whole Blood, Serum, Plasma, or Urine

Sample Size

95 µl

Detection Range

Blood Na⁺: 40-200 mmol/L

K⁺: 1.7-15 mmol/L

Cl⁻: 50-200 mmol/L

Ca²⁺: 0.3-5.0 mmol/L

Li⁺: 0.2-5.5 mmol/L

Urine Na⁺: 3-300 mmol/L

K⁺: 5-120* mmol/L

Cl⁻: 15-300 mmol/L

*60-120 requires additional dilution

Display Resolution

Na⁺: 1 mmol/L or 0.1 mmol/L

K⁺: 0.1 mmol/L or 0.01 mmol/L

Cl⁻: 1 mmol/L or 0.1 mmol/L

iCa²⁺: 0.01 mmol/L or 0.001 mmol/L

Li⁺: 0.01 mmol/L or 0.001 mmol/L

Analysis Time

60 sec without printout (60 per hour)

80 sec with printout (45 per hour)

Data Storage

1000 Patient Results

500 Level 1 QC Results

500 Level 2 QC Results

500 Level 3 QC Results

Calibration

2 point every 4 hours

1 Point after every Measurement

Reproducibility

Within Run (n=3)

Blood, Serum, Plasma

Na⁺: CV ≤ 1% (120-160 mmol/L)

K⁺: CV ≤ 2% (2.5-6 mmol/L)

Cl⁻: CV ≤ 2% (85-130 mmol/L)

iCa²⁺: SD ≤ 0.02 mmol/L (0.8 – 1.5 mmol/L)

Li⁺: SD ≤ 0.02 mmol/L (0.4 – 1.3 mmol/L)

Urine

Na⁺: CV ≤ 5% (100-250 mmol/L)

K⁺: CV ≤ 5% (10-60 mmol/L)

Cl⁻: CV ≤ 5% (100-250 mmol/L)

(Calcium and Lithium are not typically measured in urine samples)

User Input

Keypad, External Keyboard,

External Barcode Scanner (optional)

Output

16x2 place alphanumeric display

16-character width printer

RS232C serial port

Ambient Conditions

15-32°C, Room temperature (60-90°F)

<85% humidity

Languages

English, Spanish, French, German,

Italian, Portuguese

Power

100-240V ~ 50/60 Hz (self adjusting)

1.6 A max, 50 Watts max

Size & Weight

13.2 x 12.4 x 12 inches or 335 x 315 x 295 mm

14 lbs or 6 Kgs

This analyzer is produced and distributed by Diamond Diagnostics Inc. There is no affiliation, connection, sponsorship or association between Diamond Diagnostics Inc. and Roche Diagnostics or Idexx Laboratories.



DIAMOND®
DIAGNOSTICS

333 Fiske Street, Holliston, MA 01746

Tel: 508-429-0450 | Fax: 508-429-0452

www.diamonddiagnostics.com

SOP05-0021F Rev 1 11/7/11